

1. Do you have dry air (Dewpoint of -20°C)?

YES



2. Do you have air filtered to 1 micron?

YES



3. Do you have air filtered to .01 micron?

YES



4. Do you have flow meters?

NO



Customer Supplied Dryer
for -20°C and filter to .01 µm

NO



Plant Air Panel (VD11851-1010)

Include:

- Flowmeters
- Coarse Filter (5 micron)
- Coarse Coalescent Filter (.01 µm)
- Mounting Plate
- Interconnecting Tubing

Instrument Air Panel (VC11415-190)

Include:

- Flowmeters
- Coarse Filter (5 micron)
- Coarse Coalescent Filter (.01 µm)
- Mounting Plate
- Interconnecting Tubing

NO



V401-KIT with Coarse Filters
V401-4
V401-42


Include:

- Flowmeters
- Coarse Filter (5 micron)
- Coarse Coalescent Filter (.01 µm)
- Coalescent Filter (.003 µm)

Does not Include:

- Mounting Plate
- Interconnecting Tubing

NO



V401-KIT

Includes:

- Flowmeters
- Coalescent Filter (.003 µm)

Does Not Include:

- Mounting Plate
- Interconnecting Tubing